

Table of Contents: ARI Job AYO6

Client: Apex Labs

Project: A6C1076

	Page From:	Page To:
Inventory Sheet		
Cover Letter	<u>1</u>	<u>1</u>
Chain of Custody Documentation	<u>2</u>	<u>4</u>
Case Narrative, Data Qualifiers, Control Limits	<u>7</u>	<u>13</u>
General Chemistry Analysis		
Report and Summary QC Forms	<u>14</u>	<u>29</u>
General Chemistry Raw Data		
Analyst Notes and Raw Data	<u>30</u>	<u>42</u>

AV
Signature

April-15-2016
Date



Analytical Resources, Incorporated
Analytical Chemists and Consultants

April 19, 2016

Philip Nerenberg
Apex Laboratories
12232 SW Garden Place
Tigard, OR 97223

RE: Project: A6C1076
ARI Job No.: AYO6

Dear Mr. Nerenberg:

Please find enclosed the original Chain of Custody records (COCs), sample receipt documentation, and the final data for the samples from the project referenced above.

Sample receipt information and analytical details are addressed in the Case Narrative.

An electronic copy of this report and all supporting raw data will be kept on file at ARI. Should you have any questions or concerns, please feel free to call me at your convenience.

Respectfully,
ANALYTICAL RESOURCES, INC.

A handwritten signature in black ink, appearing to read "Cheronne Oreiro".

Cheronne Oreiro
Project Manager
(206) 695-6214
cheronneo@arilabs.com
www.arilabs.com

cc: eFile: AYO6

Enclosures

Chain of Custody Documentation

ARI Job ID: AYO6

SUBCONTRACT ORDER

Ayob

Apex Laboratories
A6C1076

LD
3/31/16

SENDING LABORATORY:

Apex Laboratories
12232 S.W. Garden Place
Tigard, OR 97223
Phone: (503) 718-2323
Fax: (503) 718-0333
Project Manager: Philip Nerenberg

RECEIVING LABORATORY:

Analytical Resources, INC
4611 S. 134th Place
Tukwila, WA 98168
Phone: (206) 695-6200
Fax: (206) 695-6201

Sample Name:	Sedimen	Sampled:	Sediment 0 to 6 bgs
5237-160328-DC-SED063		03/28/16 10:30	(A6C1076-02)
Analysis	Due	Expires	Comments
Sulfide, Total by PSEP (376.2) (SUB)	04/11/16 17:00	04/11/16 10:30	
<i>Containers Supplied:</i> (G)4 oz Glass Jar			

Sample Name:	Sedimen	Sampled:	Sediment 0 to 6 bgs
5237-160328-DC-SED065		03/28/16 11:00	(A6C1076-04)
Analysis	Due	Expires	Comments
Sulfide, Total by PSEP (376.2) (SUB)	04/11/16 17:00	04/11/16 11:00	
<i>Containers Supplied:</i> (G)4 oz Glass Jar			

Sample Name:	Sedimen	Sampled:	Sediment 0 to 6 bgs
5237-160328-DC-SED068		03/28/16 11:30	(A6C1076-06)
Analysis	Due	Expires	Comments
Sulfide, Total by PSEP (376.2) (SUB)	04/11/16 17:00	04/11/16 11:30	
<i>Containers Supplied:</i> (E)4 oz Glass Jar			

Sample Name:	Sedimen	Sampled:	Sediment 0 to 6 bgs
5237-160328-DC-SED070		03/28/16 12:05	(A6C1076-08)
Analysis	Due	Expires	Comments
Sulfide, Total by PSEP (376.2) (SUB)	04/11/16 17:00	04/11/16 12:05	
<i>Containers Supplied:</i> (E)4 oz Glass Jar			

Standard TAT

Level IV
DP

Susanhead

3/31/16

UPS (Shipper)

Released By	Date	Received By	Date
	3/31/16		4-1-16 01300
Released By	Date	Received By	Date

SUBCONTRACT ORDER

Ayob

Apex Laboratories
A6C1076

Sediment 0 to 6 bgs			
Sample Name:	Sedimen	Sampled:	(A6C1076-10)
Analysis	Due	Expires	Comments
5237-160328-DC-SED072		03/28/16 12:30	
Sulfide, Total by PSEP (376.2) (SUB)	04/11/16 17:00	04/11/16 12:30	
<i>Containers Supplied:</i> (E)4 oz Glass Jar			



Sediment 0 to 6 bgs			
Sample Name:	Sedimen	Sampled:	(A6C1076-12)
Analysis	Due	Expires	Comments
5237-160328-DC-SED075		03/28/16 12:50	
Sulfide, Total by PSEP (376.2) (SUB)	04/11/16 17:00	04/11/16 12:50	
<i>Containers Supplied:</i> (E)4 oz Glass Jar			

Sediment 0 to 6 bgs			
Sample Name:	Sedimen	Sampled:	(A6C1076-14)
Analysis	Due	Expires	Comments
5237-160328-DC-SED077		03/28/16 12:50	
Sulfide, Total by PSEP (376.2) (SUB)	04/11/16 17:00	04/11/16 12:50	
<i>Containers Supplied:</i> (E)4 oz Glass Jar			

Sediment 0 to 6 bgs			
Sample Name:	Sedimen	Sampled:	(A6C1076-16)
Analysis	Due	Expires	Comments
5237-160328-DC-SED077D		03/28/16 13:15	
Sulfide, Total by PSEP (376.2) (SUB)	04/11/16 17:00	04/11/16 13:15	
<i>Containers Supplied:</i> (E)4 oz Glass Jar			

Sediment 0 to 6 bgs			
Sample Name:	Sedimen	Sampled:	(A6C1076-18)
Analysis	Due	Expires	Comments
5237-160328-DC-SED082		03/28/16 13:45	
Sulfide, Total by PSEP (376.2) (SUB)	04/11/16 17:00	04/11/16 13:45	
<i>Containers Supplied:</i> (E)4 oz Glass Jar			

Sediment 0 to 6 bgs			
Sample Name:	Sedimen	Sampled:	(A6C1076-20)
Analysis	Due	Expires	Comments
5237-160328-DC-SED085		03/28/16 14:15	
Sulfide, Total by PSEP (376.2) (SUB)	04/11/16 17:00	04/11/16 14:15	
<i>Containers Supplied:</i> (E)4 oz Glass Jar			

Released By:  Date: 3/31/16
 Received By:  Date: 4-1-16 @ 1300
 Released By:  Date: _____
 Received By: _____ Date: _____


SUBCONTRACT ORDER

Apex Laboratories

A6C1076

AY06

Analysis	Due	Expires	Comments
Sample Name: 5237-160328-DC-SED087			Sediment 0 to 6 bgs Sampled: 03/28/16 14:45 (A6C1076-22)
Sulfide, Total by PSEP (376.2) (SUB)	04/11/16 17:00	04/11/16 14:45	
<i>Containers Supplied:</i> (G)4 oz Glass Jar			


3/31/16

UPS (Shipper)

Released By _____ Date _____

Received By _____ Date _____

UPS (Shipper)

 4-1-16 1300

Released By _____ Date _____

Received By _____ Date _____



Cooler Receipt Form

ARI Client: Apex
 COC No(s): _____ (NA)
 Assigned ARI Job No: AY06
 Preliminary Examination Phase:

Project Name: _____
 Delivered by: Fed-Ex UPS Courier Hand Delivered Other: _____
 Tracking No: 1Z X4720R139506 1527 NA

Were intact, properly signed and dated custody seals attached to the outside of to cooler? YES NO
 Were custody papers included with the cooler? YES NO
 Were custody papers properly filled out (ink, signed, etc.) YES NO

Temperature of Cooler(s) (°C) (recommended 2.0-6.0 °C for chemistry)
 Time: 1.4

If cooler temperature is out of compliance fill out form 00070F
 Temp Gun ID#: D005276

Cooler Accepted by: TR Date: 4-1-16 Time: 1300

Complete custody forms and attach all shipping documents

Log-In Phase:

Was a temperature blank included in the cooler? YES NO
 What kind of packing material was used? ... Bubble Wrap Wet Ice Gel Packs Baggies Foam Block Paper Other: _____
 Was sufficient ice used (if appropriate)? NA YES NO
 Were all bottles sealed in individual plastic bags? YES NO
 Did all bottles arrive in good condition (unbroken)? YES NO
 Were all bottle labels complete and legible? YES NO
 Did the number of containers listed on COC match with the number of containers received? YES NO
 Did all bottle labels and tags agree with custody papers? YES NO
 Were all bottles used correct for the requested analyses? YES NO
 Do any of the analyses (bottles) require preservation? (attach preservation sheet, excluding VOCs)... NA YES NO
 Were all VOC vials free of air bubbles? NA YES NO
 Was sufficient amount of sample sent in each bottle? YES NO
 Date VOC Trip Blank was made at ARI NA
 Was Sample Split by ARI : YES Date/Time: _____ Equipment: _____ Split by: _____

Samples Logged by: TR Date: 4-1-16 Time: 1454

**** Notify Project Manager of discrepancies or concerns ****

Sample ID on Bottle	Sample ID on COC	Sample ID on Bottle	Sample ID on COC

Additional Notes, Discrepancies, & Resolutions:

By: _____ Date: _____

<p>Small Air Bubbles = 2mm</p>	<p>Peabubbles 2-4 mm</p>	<p>LARGE Air Bubbles > 4 mm</p>	Small → "sm" (< 2 mm) Peabubbles → "pb" (2 to < 4 mm) Large → "lg" (4 to < 6 mm) Headspace → "hs" (> 6 mm)
------------------------------------	------------------------------	--	---

Case Narrative, Data Qualifiers, Control Limits

ARI Job ID: AYO6



Case Narrative

Client: Apex Laboratories
Project: A6C1076
ARI Job No.: AYO6

Sample Receipt

Analytical Resources, Inc. (ARI) accepted eleven soil samples on April 1, 2016 under ARI job AYO6. The cooler temperature measured by IR thermometer following ARI SOP was 1.4°C. For further details regarding sample receipt, please refer to the Cooler Receipt Form.

The samples were analyzed for parameters as requested.

Sulfide by SM4500-S2

Sulfide sample volumes were preserved upon receipt.

The matrix spike percent recovery of sulfide fell outside the control limits low for sample **5237-160328-DC-SED063**. All other quality control parameters were met for this analysis. No corrective action was taken.

Sample ID Cross Reference Report



ARI Job No: AYO6
Client: Apex Labs
Project Event: A6C1076
Project Name: N/A

Sample ID	ARI Lab ID	ARI LIMS ID	Matrix	Sample Date/Time	VTSR
1. 5237-160328-DC-SED063	AYO6A	16-5365	Sediment	03/28/16 10:30	04/01/16 13:00
2. 5237-160328-DC-SED065	AYO6B	16-5366	Sediment	03/28/16 11:00	04/01/16 13:00
3. 5237-160328-DC-SED068	AYO6C	16-5367	Sediment	03/28/16 11:30	04/01/16 13:00
4. 5237-160328-DC-SED070	AYO6D	16-5368	Sediment	03/28/16 12:05	04/01/16 13:00
5. 5237-160328-DC-SED072	AYO6E	16-5369	Sediment	03/28/16 12:30	04/01/16 13:00
6. 5237-160328-DC-SED075	AYO6F	16-5370	Sediment	03/28/16 12:50	04/01/16 13:00
7. 5237-160328-DC-SED077	AYO6G	16-5371	Sediment	03/28/16 12:50	04/01/16 13:00
8. 5237-160328-DC-SED077D	AYO6H	16-5372	Sediment	03/28/16 13:15	04/01/16 13:00
9. 5237-160328-DC-SED082	AYO6I	16-5373	Sediment	03/28/16 13:45	04/01/16 13:00
10. 5237-160328-DC-SED085	AYO6J	16-5374	Sediment	03/28/16 14:15	04/01/16 13:00
11. 5237-160328-DC-SED087	AYO6K	16-5375	Sediment	03/28/16 14:45	04/01/16 13:00



Analytical Resources,
Incorporated
Analytical Chemists and
Consultants

Data Reporting Qualifiers

Effective 12/31/13

Inorganic Data

- U Indicates that the target analyte was not detected at the reported concentration
- * Duplicate RPD is not within established control limits
- B Reported value is less than the CRDL but \geq the Reporting Limit
- N Matrix Spike recovery not within established control limits
- NA Not Applicable, analyte not spiked
- H The natural concentration of the spiked element is so much greater than the concentration spiked that an accurate determination of spike recovery is not possible
- L Analyte concentration is ≤ 5 times the Reporting Limit and the replicate control limit defaults to ± 1 RL instead of the normal 20% RPD

Organic Data

- U Indicates that the target analyte was not detected at the reported concentration
- * Flagged value is not within established control limits
- B Analyte detected in an associated Method Blank at a concentration greater than one-half of ARI's Reporting Limit or 5% of the regulatory limit or 5% of the analyte concentration in the sample.
- J Estimated concentration when the value is less than ARI's established reporting limits
- D The spiked compound was not detected due to sample extract dilution
- E Estimated concentration calculated for an analyte response above the valid instrument calibration range. A dilution is required to obtain an accurate quantification of the analyte.



**Analytical Resources,
Incorporated**
Analytical Chemists and
Consultants

- Q Indicates a detected analyte with an initial or continuing calibration that does not meet established acceptance criteria (<20%RSD, <20%Drift or minimum RRF).
- S Indicates an analyte response that has saturated the detector. The calculated concentration is not valid; a dilution is required to obtain valid quantification of the analyte
- NA The flagged analyte was not analyzed for
- NR Spiked compound recovery is not reported due to chromatographic interference
- NS The flagged analyte was not spiked into the sample
- M Estimated value for an analyte detected and confirmed by an analyst but with low spectral match parameters. This flag is used only for GC-MS analyses
- N The analysis indicates the presence of an analyte for which there is presumptive evidence to make a "tentative identification"
- Y The analyte is not detected at or above the reported concentration. The reporting limit is raised due to chromatographic interference. The Y flag is equivalent to the U flag with a raised reporting limit.
- EMPC Estimated Maximum Possible Concentration (EMPC) defined in EPA Statement of Work DLM02.2 as a value "calculated for 2,3,7,8-substituted isomers for which the quantitation and /or confirmation ion(s) has signal to noise in excess of 2.5, but does not meet identification criteria" (**Dioxin/Furan analysis only**)
- C The analyte was positively identified on only one of two chromatographic columns. Chromatographic interference prevented a positive identification on the second column
- P The analyte was detected on both chromatographic columns but the quantified values differ by $\geq 40\%$ RPD with no obvious chromatographic interference
- X Analyte signal includes interference from polychlorinated diphenyl ethers. (**Dioxin/Furan analysis only**)
- Z Analyte signal includes interference from the sample matrix or perfluorokerosene ions. (**Dioxin/Furan analysis only**)



**Analytical Resources,
Incorporated**
Analytical Chemists and
Consultants

Geotechnical Data

- A The total of all fines fractions. This flag is used to report total fines when only sieve analysis is requested and balances total grain size with sample weight.
- F Samples were frozen prior to particle size determination
- SM Sample matrix was not appropriate for the requested analysis. This normally refers to samples contaminated with an organic product that interferes with the sieving process and/or moisture content, porosity and saturation calculations
- SS Sample did not contain the proportion of "fines" required to perform the pipette portion of the grain size analysis
- W Weight of sample in some pipette aliquots was below the level required for accurate weighting

Analytical Method Information

Printed: 04/15/2016 7:26 am

Sulfide, SM 4500-S2 D-0, Solid (PSEP) in Solid (SM 4500-S2 D-00)

Preservation: ZnOAc, Cool <6°C

Container: Glass WM, Clear, 2 oz

Amount Required: 100 g

Hold Time: 7 days

Analyte	MDL	Reporting Limit	Surrogate %Rec	Duplicate RPD	----Matrix Spike---- %Rec	RPD	--Blank Spike / LCS-- %Rec	RPD
Sulfide	0.0750	0.500 mg/kg		20	75-125		75-125	20

General Chemistry Analysis
Report and Summary QC Forms

ARI Job ID: AYO6

SAMPLE RESULTS-CONVENTIONALS
AY06-Apex Labs



Matrix: Sediment
Data Release Authorized: W
Reported: 04/13/16

Project: NA
Event: A6C1076
Date Sampled: 03/28/16
Date Received: 04/01/16

Client ID: 5237-160328-DC-SED063
ARI ID: 16-5365 AY06A

Analyte	Date	Method	Units	RL	Sample
Preserved Total Solids	04/05/16 040516#1	SM2540G	Percent	0.01	71.66
Sulfide	04/04/16 040416#1	SM4500-S2D	mg/kg	1.39	< 1.39 U

RL Analytical reporting limit
U Undetected at reported detection limit

SAMPLE RESULTS-CONVENTIONALS
AYO6-Apex Labs



Matrix: Sediment
Data Release Authorized: ✓
Reported: 04/13/16

Project: NA
Event: A6C1076
Date Sampled: 03/28/16
Date Received: 04/01/16

Client ID: 5237-160328-DC-SED065
ARI ID: 16-5366 AYO6B

Analyte	Date	Method	Units	RL	Sample
Preserved Total Solids	04/05/16 040516#1	SM2540G	Percent	0.01	66.64
Sulfide	04/04/16 040416#1	SM4500-S2D	mg/kg	1.47	< 1.47 U

RL Analytical reporting limit
U Undetected at reported detection limit

SAMPLE RESULTS-CONVENTIONALS
AYO6-Apex Labs



Matrix: Sediment
Data Release Authorized: ✓
Reported: 04/13/16

Project: NA
Event: A6C1076
Date Sampled: 03/28/16
Date Received: 04/01/16

Client ID: 5237-160328-DC-SED068
ARI ID: 16-5367 AYO6C

Analyte	Date	Method	Units	RL	Sample
Preserved Total Solids	04/05/16 040516#1	SM2540G	Percent	0.01	68.26
Sulfide	04/04/16 040416#1	SM4500-S2D	mg/kg	1.45	< 1.45 U

RL Analytical reporting limit
U Undetected at reported detection limit

SAMPLE RESULTS-CONVENTIONALS
AYO6-Apex Labs



Matrix: Sediment
Data Release Authorized: w
Reported: 04/13/16

Project: NA
Event: A6C1076
Date Sampled: 03/28/16
Date Received: 04/01/16

Client ID: 5237-160328-DC-SED070
ARI ID: 16-5368 AYO6D

Analyte	Date	Method	Units	RL	Sample
Preserved Total Solids	04/05/16 040516#1	SM2540G	Percent	0.01	74.86
Sulfide	04/04/16 040416#1	SM4500-S2D	mg/kg	1.33	< 1.33 U

RL Analytical reporting limit
U Undetected at reported detection limit

SAMPLE RESULTS--CONVENTIONALS
AY06-Apex Labs



Matrix: Sediment
Data Release Authorized: w
Reported: 04/13/16

Project: NA
Event: A6C1076
Date Sampled: 03/28/16
Date Received: 04/01/16

Client ID: 5237-160328-DC-SED072
ARI ID: 16-5369 AY06E

Analyte	Date	Method	Units	RL	Sample
Preserved Total Solids	04/05/16 040516#1	SM2540G	Percent	0.01	58.50
Sulfide	04/04/16 040416#1	SM4500-S2D	mg/kg	1.69	< 1.69 U

RL Analytical reporting limit
U Undetected at reported detection limit

SAMPLE RESULTS-CONVENTIONALS
AY06-Apex Labs



Matrix: Sediment
Data Release Authorized: *W*
Reported: 04/13/16

Project: NA
Event: A6C1076
Date Sampled: 03/28/16
Date Received: 04/01/16

Client ID: 5237-160328-DC-SED075
ARI ID: 16-5370 AY06F

Analyte	Date	Method	Units	RL	Sample
Preserved Total Solids	04/05/16 040516#1	SM2540G	Percent	0.01	60.62
Sulfide	04/04/16 040416#1	SM4500-S2D	mg/kg	1.63	< 1.63 U

RL Analytical reporting limit
U Undetected at reported detection limit

SAMPLE RESULTS-CONVENTIONALS
AY06-Apex Labs



Matrix: Sediment
Data Release Authorized: ✓
Reported: 04/13/16

Project: NA
Event: A6C1076
Date Sampled: 03/28/16
Date Received: 04/01/16

Client ID: 5237-160328-DC-SED077
ARI ID: 16-5371 AY06G

Analyte	Date	Method	Units	RL	Sample
Preserved Total Solids	04/05/16 040516#1	SM2540G	Percent	0.01	60.82
Sulfide	04/04/16 040416#1	SM4500-S2D	mg/kg	1.64	< 1.64 U

RL Analytical reporting limit
U Undetected at reported detection limit

SAMPLE RESULTS-CONVENTIONALS
AYO6-Apex Labs



Matrix: Sediment
Data Release Authorized: ✓
Reported: 04/13/16

Project: NA
Event: A6C1076
Date Sampled: 03/28/16
Date Received: 04/01/16

Client ID: 5237-160328-DC-SED077D
ARI ID: 16-5372 AYO6H

Analyte	Date	Method	Units	RL	Sample
Preserved Total Solids	04/05/16 040516#1	SM2540G	Percent	0.01	64.26
Sulfide	04/04/16 040416#1	SM4500-S2D	mg/kg	1.55	< 1.55 U

RL Analytical reporting limit
U Undetected at reported detection limit

SAMPLE RESULTS-CONVENTIONALS
AY06-Apex Labs



Matrix: Sediment
Data Release Authorized: ✓
Reported: 04/13/16

Project: NA
Event: A6C1076
Date Sampled: 03/28/16
Date Received: 04/01/16

Client ID: 5237-160328-DC-SED082
ARI ID: 16-5373 AY06I

Analyte	Date	Method	Units	RL	Sample
Preserved Total Solids	04/05/16 040516#1	SM2540G	Percent	0.01	61.06
Sulfide	04/04/16 040416#1	SM4500-S2D	mg/kg	1.63	< 1.63 U

RL Analytical reporting limit
U Undetected at reported detection limit

SAMPLE RESULTS-CONVENTIONALS
AY06-Apex Labs



Matrix: Sediment
Data Release Authorized:
Reported: 04/13/16

Project: NA
Event: A6C1076
Date Sampled: 03/28/16
Date Received: 04/01/16

Client ID: 5237-160328-DC-SED085
ARI ID: 16-5374 AY06J

Analyte	Date	Method	Units	RL	Sample
Preserved Total Solids	04/05/16 040516#1	SM2540G	Percent	0.01	72.91
Sulfide	04/04/16 040416#1	SM4500-S2D	mg/kg	1.35	1.54

RL Analytical reporting limit
U Undetected at reported detection limit

SAMPLE RESULTS-CONVENTIONALS
AYO6-Apex Labs



Matrix: Sediment
Data Release Authorized: W
Reported: 04/13/16

Project: NA
Event: A6C1076
Date Sampled: 03/28/16
Date Received: 04/01/16

Client ID: 5237-160328-DC-SED087
ARI ID: 16-5375 AYO6K

Analyte	Date	Method	Units	RL	Sample
Preserved Total Solids	04/05/16 040516#1	SM2540G	Percent	0.01	69.85
Sulfide	04/04/16 040416#1	SM4500-S2D	mg/kg	1.43	10.8

RL Analytical reporting limit
U Undetected at reported detection limit

MS/MSD RESULTS-CONVENTIONALS
AY06-Apex Labs



Matrix: Sediment
Data Release Authorized: w
Reported: 04/13/16

Project: NA
Event: A6C1076
Date Sampled: 03/28/16
Date Received: 04/01/16

Analyte	Date	Units	Sample	Spike	Spike Added	Recovery
---------	------	-------	--------	-------	-------------	----------

ARI ID: AY06A Client ID: 5237-160328-DC-SED063

Sulfide	04/04/16	mg/kg	< 1.39	61.0	197	31.0%
---------	----------	-------	--------	------	-----	-------

REPLICATE RESULTS-CONVENTIONALS
AY06-Apex Labs



Matrix: Sediment
Data Release Authorized: *W*
Reported: 04/13/16

Project: NA
Event: A6C1076
Date Sampled: 03/28/16
Date Received: 04/01/16

Analyte	Date	Units	Sample	Replicate (s)	RPD/RSD
ARI ID: AY06A Client ID: 5237-160328-DC-SED063					
Preserved Total Solids	04/05/16	Percent	71.66	77.15	7.4%
Sulfide	04/04/16	mg/kg	< 1.39	< 1.39	NA

LAB CONTROL RESULTS-CONVENTIONALS
AY06-Apex Labs



Matrix: Sediment
Data Release Authorized: ✓
Reported: 04/13/16

Project: NA
Event: A6C1076
Date Sampled: NA
Date Received: NA

Analyte/Method	QC ID	Date	Units	LCS	Spike Added	Recovery
Sulfide SM4500-S2D	PREP	04/04/16	mg/kg	7.19	7.19	100.0%

METHOD BLANK RESULTS-CONVENTIONALS
AY06-Apex Labs



Matrix: Sediment
Data Release Authorized: w
Reported: 04/13/16

Project: NA
Event: A6C1076
Date Sampled: NA
Date Received: NA

Analyte	Date	Units	Blank	QC ID
Preserved Total Solids	04/05/16	Percent	< 0.01 U	ICB
Sulfide	04/04/16	mg/kg	< 0.05 U	PREP

General Chemistry Raw Data
Analyst Notes and Raw Data

ARI Job ID: AYO6

h-6-16

TOTAL SOLIDS/VOLATILE SOLIDS (TS / TVS) BENCHSHEET
 (dry at 104 (12-24 hr) then combust at 550 (30 min))

SOLIDS DATE: 4/5/16 (A) ANALYST: KE 12:38
 Instrumentation ZnOAc PRES. Analytical Balance: 1123230597

Drying Overs: 12 Muffle Furnace: N/A
 TS (%) calculated as:
 Final dry wt (g) = (Dry Wt - Tare Wt)
 TS = (Final Dry Wt)/(grams Sample-Tare)

TVS (mg/kg dry wt) calculated as:
 Final ash wt (g) = (min ash wt - tare wt)
 TVS (mg/kg) = [(Dry wt-Ash wt)/(dry weight)] * 1,000,000
 if ash wt > dry wt, "Chk for Err"
 if dry wt-ash wt < 0.001 g, "% (1/dry wt) * 1,000,000"

SAMPLE ID	DISH #	SAMPLE (grams)	TARE WT (grams)	DRY WT 104C (grams)	dry wt (g)	TS (%)	ASH WT 550C (grams)		Ash Wt (g)	TVS (mg/kg) (%)
							1	2		
Blank			1.1247	1	0.00					
AYO5 A1		6.6143	1.1569	4.9358	3.78	69.24%				
AYO5 A1 dup		6.5521	1.1506	4.9538	3.80	70.41%				

RPD = 1.67%

AYO5 B1		6.9683	1.1284	5.3012	4.17	71.45%				
AYO5 C1		6.4261	1.1497	4.9768	3.83	72.53%				
AYO5 D1		6.5631	1.0949	5.1167	4.02	73.55%				
AYO5 E1		6.5198	1.1118	5.0764	3.96	73.31%				
AYO5 F1		6.8561	1.0984	5.3961	4.30	74.64%				
AYO5 G1		6.5329	1.0896	4.9555	3.87	71.02%				
AYO5 H1		6.9427	1.1232	5.1559	4.03	69.30%				
AYO6 A1		6.4126	1.0854	4.9029	3.82	71.66%				
AYO6 A1 dup		6.8207	1.1336	5.5211	4.39	77.15%				

RPD = 7.38%

AYO6 B1		6.8375	1.1451	4.9383	3.79	66.64%				
AYO6 C1		6.7209	1.0803	4.9308	3.85	68.26%				
AYO6 D1		7.4518	1.1438	5.8661	4.72	74.86%				
AYO6 E.1		7.2274	1.1525	4.7065	3.55	58.50%				
AYO6 F1		6.4437	1.1541	4.3608	3.21	60.62%				
AYO6 G1		7.7895	1.1270	5.1793	4.05	60.82%				
AYO6 H1		7.7771	1.1496	5.4082	4.26	64.26%				
AYO6 I 1		7.6391	1.0944	5.0906	4.00	61.06%				

W
4-12-16

SULFIDE BENCHSHEET (SM 4500-S2- D. Methylene Blue)		Date Time	Analyst	
Soils, sediments and solid phase samples		4/4/16 11:34	RLM	
Distillation Finish		4/6/2016 16:26 AM	NN	
If distilled, specify Procedure: PSEP		ZnOAc: D004538	Balance: 19350128	
1. Standardization of sodium thiosulfate titrant		Buret used for titrations: S2		
Thiosulfate ID: D004645		Titration of bi-iodate with thiosulfate		
Bi-iodate ID: D003541				
Stock bi-iodate = 0.8118 grams to 1000 mL		ml bi-iodate =	3.00	3.00
Normality = 0.025		ml thiosulfate =	3.14	3.13
Normality thiosulfate = (mL bi-iodate*normality) / mL thiosulfate =			0.024	0.024
			0.024	0.024
2. Normality of Iodine		Titration of Iodine with thiosulfate		
Iodine ID: E000905		mL Iodine =	3.00	3.00
		mL thiosulfate =	3.10	3.10
Normality iodine = (mL thiosulfate*normality) / mL iodine=			0.025	0.025
			0.024	0.025
3. Standardization of sodium sulfide stock		Titration of standard with Thiosulfate		
Stock ID = E000991		mL Standard =	1.00	1.00
Approx conc in 100ml		mL iodine =	3.00	3.00
g Na2S 0.4666 mg/mL = 0.623		mL thiosulfate =	1.22	1.20
Sulfide (mg/mL) = (((mL iodine*ni)-(mL thio *nthio))*16) / mL standard =			0.71	0.72
			0.72	0.719
Intermediate Standard		mL required for for 0.025 mg/mL 8.7		
Add	8.7	ml stk to	250	ml 0.2 N ZnOAc = 0.025 mg/mL
4. Calibration Standard Curve		spectrophotometer used: UV 1800-1		
Inter Std Volume (mL)	Final Volume (mL)	Calc Conc (mg S/L)	Absorbance @650 nm	
			1	2
0.00	50	0.000	0.000	0.000
0.10	50	0.050	0.034	0.034
0.25	50	0.125	0.091	0.091
0.50	50	0.250	0.174	0.174
1.00	50	0.500	0.331	0.331
2.00	50	1.000	0.660	0.660
Calib Verif Std =		1.0 ml INT to	50	ml ZnOAc=
Distillation Std =		1.0 ml stk to	100	=
				0.500 mg/l
				7.19 mg/l

SAMPLE DATA

enter dilution as mL final/mL sample

SAMPLE ID	Distillation Data			Spectrophotometric Data			SAMPLE DATA	
	SAMPLE SIZE	% Solids	TRAP VOLUME (ml)	Dilution Factor	Abs @ 650 nm	regressed Conc (mg S/L)	CORR CONC (ppm)	
					Sample	Bkg		
ICB		na	na	1.00	0.003		-0.002	< 0.05 OK!
ICV		na	na	1.00	0.329		0.495	0.495 98.89%
Distilled samples								
Dist Blk	100	100%	100	1.00	0.004		0.000	< 0.05 OK!
Dist Chk	100	100%	100	10.00	0.476		0.719	7.186 99.99%
Soil Samples								
	(grams)	% Solids	(mL)		Sample	Bkg	(mg/L)	mg/kg
AYO6 A1	5.014	71.66%	100	1.00	0.002		-0.004	< 1.392
AYO6 A1 dup	5.031	71.66%	100	1.00	0.001		-0.005	< 1.388 NA
AYO6 A1 ms	5.098	71.66%	100	10.00	0.151		0.223	61.171 31.09%
	Spike at		1.00	ml stock to	3.653	g dry wt =		196.738 mg/kg
AYO6 B1	5.091	66.64%	100	1.00	-0.002		-0.010	< 1.474
AYO6 C1	5.049	68.26%	100	1.00	-0.002		-0.010	< 1.451
AYO6 D1	5.015	74.86%	100	1.00	-0.004		-0.013	< 1.332
AYO6 E1	5.055	58.50%	100	1.00	-0.003		-0.011	< 1.692
AYO6 F1	5.053	60.62%	100	1.00	-0.007		-0.017	< 1.633
Cal Blk		na	na	1.00	-0.028		-0.049	< 0.05 OK!
CCV		na	na	1.00	0.302		0.454	0.454 90.67%
AYO6 G1	5.016	60.82%	100	1.00	0.005		0.001	< 1.64
AYO6 H1	5.014	64.26%	100	1.00	0.004		0.000	< 1.553
AYO6 I1	5.013	61.06%	100	1.00	0.007		0.004	< 1.634
AYO6 J1	5.087	72.91%	100	1.00	0.042		0.057	1.548

SAMPLE DATA

enter dilution as mL final/mL sample

SAMPLE ID	Distillation Data			Spectrophotometric Data			SAMPLE DATA		
	SAMPLE SIZE	% Solids	TRAP VOLUME (ml)	Dilution Factor	Abs @ 650 nm		regressed Conc (mg S/L)	CORR CONC (ppm)	
					Sample	Bkg			
AYO6 K1	5.019	69.85%	100	1.00	0.253		0.379	10.807	
AYO5 A1	5.025	69.24%	100	1.00	0.004		0.000	< 1.438	
AYO5 B1	5.027	71.45%	100	1.00	0.033		0.044	< 1.393	
AYO5 C1	5.003	72.53%	100	1.00	0.000		-0.007	< 1.379	
AYO5 D1	5.019	73.55%	100	1.00	-0.002		-0.010	< 1.355	
AYO5 E1	5.091	73.31%	100	1.00	-0.004		-0.013	< 1.34	
Cal Blk		na	na	1.00	-0.023		-0.042	< 0.05	OK!
CCV		na	na	1.00	0.313		0.470	0.470	94.02%
AYO5 F1	5.007	74.64%	100	1.00	0.004		0.000	< 1.339	
AYO5 G1	5.016	71.02%	100	1.00	0.003		-0.002	< 1.404	
AYO5 H1	5.035	69.30%	100	1.00	0.001		-0.005	< 1.434	
AYO6 A1 ms	5.098	71.66%	100	10.00	0.154		0.228	62.422	31.73%
		<i>Spike at</i>	<i>1.00</i>	<i>ml stock to</i>	<i>3.653</i>	<i>g dry wt =</i>	<i>196.738 mg/kg</i>		
Cal Blk		na	na	1.00	-0.008		-0.019	< 0.05	OK!
CCV		na	na	1.00	0.348		0.524	0.524	104.68%

Conventionals Distillation and Digestion Log



Method:

Sulfide, PSEP, Solid

Matrix:

Soil

Analyst: RLM

Prep Date, Time: 4/4/16 11:34

Sample Preparation Log

Reagents, Equipment

HCl+Al: D604196
 pH Indicator: D603923
 0.2N ZnOAc: D604538
 Balance: 19350128
 Sulfide Stock: E000991

Step By Step

5 grams sample
 Fill traps to line with 0.2N ZnOAc
 LCS and MS get 1mL S2 Stock
 100mL dispersing water
 ~5mL HCl+Al to pH < 3 by
 Bromophenol Blue Indicator
 60 minutes at 90C
 Decant to 100 mL with ZnOAc

Lab Number	Name	Sample Amount (g/mL)	Final Volume (mL)	Spike ID	Spike Volume (µL)
PB			100		
LCS				E000991	1000
Ay06	A1	5.014			
	Alup	5.031			
	Alms	5.098		E000991	1000
	B1	5.091			
	C1	5.049			
	D1	5.015			
	E1	5.050			
	F1	5.083			
	G1	5.016			
	H1	5.014			
	I1	5.013			
	J1	5.087			
	K1	5.019			
Ay05	A1	5.025			
	B1	5.027			
	C1	5.003			
	D1	5.019			
	E1	5.091			
	F1	5.007			
	G1	5.016			
	H1	5.035			

RLM
4/4/16

Revision: 0005 7/1/2015

SULFIDE TITRATION

Buret used for titrations: S2

Standardization of sodium thiosulfate titrant

Thiosulfate ID: D004645 Analyst: CAL / NAN
 Bi-iodate ID: ~~E002341~~ D003541 Date & Time: 4-6-16
 Stock bi-iodate = 0.8118 grams to 1000 mL
 Normality =
 Titration of bi-iodate with thiosulfate

mL bi-iodate =	3.00	3.00	3.00	
mL thiosulfate =	3.14	3.13	3.14	nthio
Normality thiosulfate =				

(mL bi-iodate * normbio) / mL thiosulfate

Normality of Iodine

Iodine ID: E000905 Analyst:
 Titration of iodine with thiosulfate Date & Time:
 mL iodine =

3.00	3.00	3.00
------	------	------

 mL thiosulfate =

3.10	3.10	3.07
------	------	------

 ni
 Normality iodine =

--	--	--

 (mL thiosulfate * nthio) / mL iodine

Standardization of Sodium Sulfide Stock

Stock ID = E000991 Analyst:
 Approx conc in 60 mL Date & Time:
 g Na₂S = 0.4666 mg/mL =
 Titration of standard with thiosulfate

mL Standard =	1.00	1.00	1.00	
mL iodine =	3.00	3.00	3.00	
mL thiosulfate =	1.22	1.20	1.21	stkconc (mg/mL)
Sulfide (mg/mL) =				

{[(mL iodine * ni) - (mL thio * nthio)] * 16} / mL standard
 mL required for for 0.025 mg/mL

Standardization of Sodium Sulfide Stock


Stock ID = Analyst:
 Approx conc in 60 mL Date & Time:
 g Na₂S = mg/mL =
 Titration of standard with thiosulfate

mL Standard =	1.00	1.00	1.00	
mL iodine =	3.00	3.00	3.00	
mL thiosulfate =				stkconc (mg/mL)
Sulfide (mg/mL) =				

{[(mL iodine * ni) - (mL thio * nthio)] * 16} / mL standard
 mL required for for 0.025 mg/mL

Quantitative Measurement Report

Data set: Z:\Shimadzu Spec Methods\SULFIDE DATA 2016\Sulfide 040616b nn.pho



Software Information

Software Name: UVPProbe
Version: 2.51
Mode: Security Mode

Data Information

Filename: Z:\Shimadzu Spec Methods\SULFIDE DATA 2016
\Sulfide 040616b nn.pho
Title: NN APD Sulfide 04 06 16
Analyst: Nhan Nguyen
Date/Time: 04/11/2016 12:16:51 PM
Comments: added dilutions

Instrument Information

Instrument Name: CONV-UV-1
Instrument Type: UV-1800 Series
Model (S/N): 206-25400-42 (A11455350830)

$x = 1.52269 y - 0.00631642$
Correlation Coefficient $r^2 = 0.99971$

Standard Table

	Sample ID	Date	Time	Conc	Abs@650.0	Comments
1	Std 1 (Zero)	04/06/201	02:58:03 PM	0.000	-0.000	
2	Std 2 (0.10 mL)	04/06/201	02:58:29 PM	0.050	0.034	
3	Std 3 (0.25 mL)	04/06/201	02:58:50 PM	0.125	0.091	
4	Std 4 (0.50 mL)	04/06/201	02:59:15 PM	0.250	0.174	
5	Std 5 (1.00 mL)	04/06/201	02:59:38 PM	0.500	0.331	
6	Std 6 (2.00 mL)	04/06/201	03:00:04 PM	1.000	0.660	
7						

Quantitative Measurement Report

Data set: Z:\Shimadzu Spec Methods\SULFIDE DATA 2016\Sulfide 040616b nn.pho

Sample Table

	Sample ID	Date	Time	Conc	Abs@650.0	DF	Comments
1	ICB	04/06/2016	03:01:10 PM	-0.012	-0.004	1.000	
2	ICV	04/06/2016	03:01:42 PM	0.505	0.336	1.000	
3	AYN4 A1	04/06/2016	03:20:22 PM	0.022	0.019	1.000	
4	AYN4 A1 Dup	04/06/2016	03:21:07 PM	0.024	0.020	1.000	
5	AYN4 A1 MS	04/06/2016	03:21:48 PM	0.349	0.233	1.000	
6	AYN4 B1	04/06/2016	03:24:12 PM	0.015	0.014	1.000	
7	AYN4 C1	04/06/2016	03:25:32 PM	-0.026	-0.013	1.000	
8	AYN4 D1	04/06/2016	03:26:18 PM	-0.026	-0.013	1.000	
9	AYN4 E1	04/06/2016	03:28:02 PM	-0.026	-0.013	1.000	
10	AYN4 A1 MS2	04/06/2016	03:42:33 PM	0.376	0.251	1.000	
11	AYN4 A1 MS3	04/06/2016	04:00:15 PM	0.382	0.255	1.000	
12	CCB	04/06/2016	04:02:47 PM	-0.013	-0.004	1.000	
13	CCV	04/06/2016	04:03:25 PM	0.475	0.316	1.000	
14	ICB 2	04/06/2016	04:26:42 PM	-0.001	0.003	1.000	
15	ICV 2	04/06/2016	04:27:12 PM	0.494	0.329	1.000	
16	PREPBLANK	04/06/2016	04:28:11 PM	-0.001	0.004	1.000	
17	LCS 2	04/06/2016	04:28:47 PM	0.719	0.476	10.00	
18	AYO6 A1	04/06/2016	04:31:06 PM	-0.004	0.002	1.000	
19	AYO6 A1 DU	04/06/2016	04:32:06 PM	-0.004	0.001	1.000	
20	AYO6 A1 MS	04/06/2016	04:32:58 PM	0.223	0.151	10.00	
21	AYO6 B1	04/06/2016	04:35:24 PM	-0.010	-0.002	1.000	
22	AYO6 C1	04/06/2016	04:36:26 PM	-0.009	-0.002	1.000	
23	AYO6 D1	04/06/2016	04:37:04 PM	-0.012	-0.004	1.000	
24	AYO6 E1	04/06/2016	04:37:51 PM	-0.012	-0.003	1.000	
25	AYO6 F1	04/06/2016	04:38:36 PM	-0.016	-0.007	1.000	
26	CCB 2	04/06/2016	04:39:08 PM	-0.049	-0.028	1.000	
27	CCV 2	04/06/2016	04:39:51 PM	0.453	0.302	1.000	
28	AYO6 G1	04/06/2016	05:05:51 PM	0.001	0.005	1.000	
29	AYO6 H1	04/06/2016	05:06:46 PM	0.000	0.004	1.000	
30	AYO6 I1	04/06/2016	05:07:32 PM	0.004	0.007	1.000	
31	AYO6 J1	04/06/2016	05:08:20 PM	0.058	0.042	1.000	
32	AYO6 K1	04/06/2016	05:09:04 PM	0.380	0.253	1.000	
33	AYO5 A1	04/06/2016	05:09:55 PM	-0.001	0.004	1.000	
34	AYO5 B1	04/06/2016	05:10:38 PM	0.044	0.033	1.000	
35	AYO5 C1	04/06/2016	05:11:33 PM	-0.007	-0.000	1.000	
36	AYO5 D1	04/06/2016	05:12:45 PM	-0.009	-0.002	1.000	
37	AYO5 E1	04/06/2016	05:13:29 PM	-0.012	-0.004	1.000	
38	CCB 3	04/06/2016	05:14:17 PM	-0.041	-0.023	1.000	
39	CCV 3	04/06/2016	05:15:20 PM	0.471	0.313	1.000	

Quantitative Measurement Report

Data set: Z:\Shimadzu Spec Methods\SULFIDE DATA 2016\Sulfide 040616b nn.pho

Sample Table

	Sample ID	Date	Time	Conc	Abs@650.0	DF	Comments
40	AYO5 F1	04/06/2016	05:33:39 PM	-0.001	0.004	1.000	
41	AYO5 G1	04/06/2016	05:34:26 PM	-0.001	0.003	1.000	
42	AYO5 H1	04/06/2016	05:35:04 PM	-0.005	0.001	1.000	
43	AYO6 A1 MS2	04/06/2016	05:36:11 PM	0.228	0.154	10.00	
44	CCB 4	04/06/2016	05:40:40 PM	-0.019	-0.008	1.000	
45	CCV 4	04/06/2016	05:41:03 PM	0.523	0.348	1.000	
46	ICB 3	04/06/2016	06:05:14 PM	-0.005	0.001	1.000	
47	ICV 3	04/06/2016	06:06:00 PM	0.528	0.351	1.000	
48	PREPBLANK	04/06/2016	06:09:18 PM	-0.008	-0.001	1.000	
49	LCS3	04/06/2016	06:09:41 PM	0.724	0.480	10.00	
50	AYO7 A1	04/06/2016	06:11:58 PM	-0.028	-0.014	1.000	
51	AYO7 A1 DU	04/06/2016	06:12:26 PM	-0.038	-0.021	1.000	
52	AYO7 A1 MS	04/06/2016	06:13:14 PM	0.565	0.375	10.00	
53	AYO7 B1	04/06/2016	06:15:46 PM	-0.027	-0.014	1.000	
54	AYO7 C1	04/06/2016	06:16:41 PM	-0.022	-0.011	1.000	
55	AYO7 D1	04/06/2016	06:17:03 PM	-0.021	-0.010	1.000	
56	AYO7 E1	04/06/2016	06:17:28 PM	-0.037	-0.020	1.000	
57	AYO7 F1	04/06/2016	06:17:52 PM	-0.022	-0.010	1.000	
58	CCB 5	04/06/2016	06:18:41 PM	-0.033	-0.018	1.000	
59	CCV 5	04/06/2016	06:19:10 PM	0.482	0.320	1.000	
60	AYO7 G1	04/06/2016	06:38:00 PM	0.004	0.007	1.000	
61	AYO7 H1	04/06/2016	06:38:26 PM	0.002	0.006	1.000	
62	AYO7 I1	04/06/2016	06:38:50 PM	0.003	0.006	1.000	
63	AYO7 J1	04/06/2016	06:39:18 PM	0.002	0.005	1.000	
64	AYR5 A1	04/06/2016	06:39:42 PM	0.003	0.006	1.000	
65	AYR5 B1	04/06/2016	06:40:06 PM	0.002	0.006	1.000	
66	AYR5 C1	04/06/2016	06:40:35 PM	0.044	0.033	1.000	
67	AYR5 D1	04/06/2016	06:41:10 PM	0.081	0.058	1.000	
68	AYR5 E1	04/06/2016	06:41:55 PM	0.000	0.004	1.000	
69	AYR6 A1	04/06/2016	06:42:17 PM	-0.001	0.003	1.000	
70	CCB 6	04/06/2016	06:42:43 PM	-0.004	0.001	1.000	
71	CCV 6	04/06/2016	06:43:05 PM	0.505	0.336	1.000	
72	AYR6 B1	04/06/2016	06:43:51 PM	0.037	0.028	1.000	
73	AYR6 C1	04/06/2016	06:44:15 PM	-0.002	0.003	1.000	
74	AYR6 D1	04/06/2016	06:44:39 PM	0.642	0.426	1.000	
75	AYR6 E1	04/06/2016	06:45:14 PM	0.548	0.364	1.000	
76	CCB 7	04/06/2016	06:45:50 PM	-0.015	-0.006	1.000	
77	CCV 7	04/06/2016	06:46:14 PM	0.509	0.338	1.000	
78	ICB 4	04/06/2016	07:05:04 PM	-0.004	0.001	1.000	

Quantitative Measurement Report

Data set: Z:\Shimadzu Spec Methods\SULFIDE DATA 2016\Sulfide 040616b nn.pho

Sample Table

	Sample ID	Date	Time	Conc	Abs@650.0	DF	Comments
79	ICV 4	04/06/2016	07:05:27 PM	0.506	0.337	1.000	
80	PREPBLANK	04/06/2016	07:06:02 PM	-0.005	0.001	1.000	
81	LCS 3	04/06/2016	07:06:26 PM	0.727	0.482	10.00	
82	AYR6 F1	04/06/2016	07:07:22 PM	0.045	0.034	1.000	
83	AYR6 F1 DUP	04/06/2016	07:07:45 PM	0.035	0.027	1.000	
84	AYRS F1 MS	04/06/2016	07:08:10 PM	0.645	0.428	10.00	
85	CCB 8	04/06/2016	07:09:45 PM	-0.014	-0.005	1.000	
86	CCV 8	04/06/2016	07:10:00 PM	0.518	0.344	1.000	
87							

Quantitative Measurement Report

Data set: Z:\Shimadzu Spec Methods\SULFIDE DATA 2016\Sulfide 040616b nn.pho

[Wavelengths]

Wavelength Name: Abs@650.0
Wavelength: 650.00 nm

[Calibration Curve]

Column for Cal. Curve: Abs@650.0
Cal. Curve Type: Multi Point
Cal. Curve Unit: mg/L
Selected Wavelength: Abs@650.0
Calibration Equation: $\text{Conc} = K1*(\text{Abs}) + K0$
Zero Interception: Not Selected

[Measurement Parameters(Standard)]

Data Acquired by: Instrument
Delay sample read: Disabled
Repeat: Disabled

[Measurement Parameters(Sample)]

Data Acquired by: Instrument
Delay sample read: Disabled
Repeat: Disabled

[Equations]

Equation Name: AdjConc
Equation: $\text{Conc} * \text{DF}$
Units: mg/L

[Pass Fail]

[Method Summary]

Title: Sulfide Colorimetry
Date/Time: 01/06/2016 05:28:50 PM
Comments:
Sample Preparations:

[Instrument Properties]

Instrument Type: UV-1800 Series
Measuring Mode: Absorbance
Slit Width: 1.0 nm
Light Source Change Wavelength: 340.0 nm
S/R Exchange: Normal

[Attachment Properties]

Attachment: None



Corrective Actions Inorganic Analyses

Criteria Flagged:

ARI Job No.: AY06

Unacceptable Blank:

Date of Event: 4-6-2016

Unacceptable Duplicate:

Client ID: _____

Unacceptable Spike:

Method/Element: Sulfide

Unacceptable Reference:

Prep Code: _____

Details of Problem/Recommended Corrective Action:

Low spike recovery @ 31.09%

Samples Affected: AY06 A'

Corrective Action Taken: Re-read matrix spike to verify; likely matrix interference. Data reported as-is.

Analyst Initials: AW

Supervisor: W

Date: 4-11-16

Date: 4-12-16